



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that ACM Global, using Tosoh G8 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: <b>G8</b>	Calibrator Lot: <b>ZS4001</b>	Column Lot: <b>F</b>
Reagent Lot: <b>F8-112Y, F8-211Y, F8-310Y, HW005F/B</b>	Calibrator Assigned Values: <b>5.84%, 10.78%</b>	

Date of Certification: August 1, 2015

Certification Expires: August 1, 2016

*Daniel Soule*  
NGSP Steering Committee Chair

*Rondie R Little PhD.*  
NGSP Network Coordinator

*Shawn Smith*  
SRL director/ supervisor